



## ATEC VeA Alignment App

### INSTRUCTIONS FOR USE

## General Information

VEA is intended for use by trained healthcare professionals as a standalone software for viewing, measuring, and storing of x-ray images. The app is a medical device software used to measure spinopelvic parameters from patient x-ray images taken with the device's camera. These measured parameters provide a quantifiable way to assess a patient's spinal deformity and correction correlated to health-related quality of life (HRQOL) scores based on published literature. VEA is intended for use on mobile phones only. Do not use VEA on any type of tablet, iPod, or non-mobile phone device. For VEA to work appropriately, users must be running minimum operating systems as defined in the Minimum Device Requirement table. Minimum device requirements are listed below.

Minimum Device Requirements	
	iOS
Minimum Screen Resolution (Pixels)	828 x 1792
Minimum Screen Size (Inches)	6.1
Rear Camera Resolution	12 MP
Operating System	iOS 16

Compatible Operating Systems	
	iOS
Operating System	iOS 16 and iOS 17

### INDICATIONS FOR USE:

VEA assists healthcare professionals in viewing, measuring, and storing images as well as planning orthopedic surgeries. The app allows the healthcare professional to perform generic and specialty measurements of the images, and to plan surgical procedures. The app includes tools for measuring anatomical components for implant selection and offers the possibility to share data among VEA users. The app is for planning use only and is not intended for primary diagnostic use.

### CONTRAINDICATIONS:

None.

### PRECAUTIONS:

- Disclaimer: "Sales representatives must adhere to HIPAA regulations by only accessing patient data relevant to their assigned care responsibilities."
- Confirm that at least two of the image corners align with the green corner markers on the image capture display. This will assist in confirming that the device is parallel to and aligned vertically with the image taken. Do not tilt or rotate the device when taking an image. Images taken with the device not properly aligned may result in inaccurate measurements. Measurements are accurate within 2 degrees, provided the device is not tilted more than 5 degrees. Take care to ensure the device is not rotated more than 5 degrees when capturing an image.
- Proper anatomical landmark placement using the features described is critical to the accuracy of the calculated parameters.

- For accurate parameter calculation it is important for the user to visually confirm the correct landmark placement on the x-ray image in the 'Results' screen and use the described back arrow icon to correct any identified errors in landmark placement.
- This application was designed to measure spinopelvic features on a vertically oriented x-ray (i.e., cranial towards top, caudal towards bottom). Measurements made outside of these conditions may be inaccurate.
- Clinical judgment and experience are required to properly use the software.
- Operating System (i.e., iOS, Android) updates are not recommended to be executed until the compatibility has been tested by Alphatec. However, security updates to the device operating system are still encouraged.
- To maintain the accuracy of the alignment results,
  - Refrain from disabling the default device auto-brightness feature. If this feature is disabled, the minimum brightness should be set to 25% or higher.
  - If the screen is not visible due to the mirroring of light or glare/reflections, adjust the angle with which the screen is viewed.

For Surgical Technique Guides or additional information regarding the products, please contact your local representative or Alphatec Spine, Inc., Customer Service directly at [customerservice@atecspine.com](mailto:customerservice@atecspine.com).

For additional information, please reference the ATEC Alignment App User Manual (part number LIT-84919).



**CAUTION: Federal law (USA) restricts these devices to sale by or on the order of a physician.**

For a listing of Symbols and Explanations, see [atecspine.com/eifu](https://atecspine.com/eifu)



**ALPHATEC SPINE, INC**  
1950 Camino Vida Roble  
Carlsbad, CA 92008, USA  
(760) 431-9286  
(800) 922-1356  
[www.atecspine.com](http://www.atecspine.com)

Alphatec Spine, ATEC Spine, the ATEC logo, EOS, and VEA are trademarks or registered trademarks of Alphatec Holdings, Inc., its affiliates and/or subsidiary companies, registered in the USA and other countries. All other trademarks belong to their respective owners. For patent information, please visit <https://atecspine.com/patent-marking/>. © 2024 Alphatec Spine, Inc. All rights reserved.